

# CE CERTIFICATE

**WE HEREBY CERTIFY THAT:**

Certificate No.: R04081604E

**EQUIPMENT:**

Linear Actuator

**Model No.:**

LA3X(0~9)

**Serial No.:**

CB0X(0~9); CB1X(0~9); CB2X(0~9); HB8X(0~5); HB3X(0~5); ACPX(0~5)

**Applicant:**

LINKAN ASIA CO.

5F, NO.78, SEC. 1, KWANG FU RD., SAN-CHUNG CITY,  
TAIPEI HSIEN TAIWAN R.O.C.

**IS IN COMPLIANCE TO THE EUROPEAN COUNCIL DIRECTIVE 89/336/EEC  
AS AMENDED BY DIRECTIVES 92/31/EEC AND 93/68/EEC WITHIN CE MARKING  
REQUIREMENT.**

**AND KEEPS ALL REQUIREMENTS ACCORDING THE FOLLOWING REGULATIONS**

EN 55011:1998: Class B	EN 61000-3-2:2001	EN 61000-3-3:2001
EN 60601-1-2 : 2001	EN 61000-4-2:1995	EN 61000-4-3:2002
	EN 61000-4-4:1995	EN 61000-4-5:1995
	EN 61000-4-6:1996	EN 61000-4-11:1994

Taiwan  
Aug. 31, 2004Matrix Test Laboratory  
2F, No. 146, Jianyi Rd.,  
Janghe City, Taipei, Taiwan  
Peter Chin, Managing Director

Specifications are subject to change without prior notice. It is the responsibility for the product users to determine the suitability of LINKAN products for a specific application. LINKAN will be responsible for replacing/ repairing defective products covered by the warranty if promptly returned to the factory. No liability is assumed beyond such replacement/ repair.

**LINKAN ASIA CO.**

2006/3/20